



CERTIFICATE OF CALIBRATION

Certificate No.: 14CA0529 01-01 Page 1 of 2

Item tested

| | | | |
|-----------------------|----------------------------|---|------------|
| Description: | Sound Level Meter (Type 1) | , | Microphone |
| Manufacturer: | B & K | , | B & K |
| Type/Model No.: | 2236 | , | 4188 |
| Serial/Equipment No.: | 2100736 | , | 2157055 |
| Adaptors used: | - | , | - |

Item submitted by

Customer Name: Lam Geotechnics Limited
Address of Customer: -
Request No.: -
Date of receipt: 29-May-2014

Date of test: 29-May-2014

Reference equipment used in the calibration

| Description: | Model: | Serial No. | Expiry Date: | Traceable to: |
|---------------------------------|----------|------------|--------------|---------------|
| Multi function sound calibrator | B&K 4226 | 2288444 | 22-Jun-2014 | CIGISMEC |
| Signal generator | DS 360 | 33873 | 09-Apr-2015 | CEPREI |
| Signal generator | DS 360 | 61227 | 09-Apr-2015 | CEPREI |

Ambient conditions

Temperature: 22 ± 1 °C
Relative humidity: 60 ± 10 %
Air pressure: 1000 ± 10 hPa

Test specifications

- 1, The Sound Level Meter has been calibrated in accordance with the requirements as specified in BS 7580: Part 1: 1997 and the lab calibration procedure SMTP004-CA-152.
- 2, The electrical tests were performed using an electrical signal substituted for the microphone which was removed and replaced by an equivalent capacitance within a tolerance of ±20%.
- 3, The acoustic calibration was performed using an B&K 4226 sound calibrator and corrections was applied for the difference between the free-field and pressure responses of the Sound Level Meter.

Test results

This is to certify that the Sound Level Meter conforms to BS 7580: Part 1: 1997 for the conditions under which the test was performed.

Details of the performed measurements are presented on page 2 of this certificate.

Actual Measurement data are documented on worksheets.

Approved Signatory:


Huang Jian Min / Feng Jun Qi

Date: 30-May-2014

Company Chop:



Comments: The results reported in this certificate refer to the condition of the instrument on the date of calibration and carry no implication regarding the long-term stability of the instrument.



CERTIFICATE OF CALIBRATION

(Continuation Page)

Certificate No.: 14CA0529 01-01 Page 2 of 2

1, Electrical Tests

The electrical tests were performed using an equivalent capacitance substituted for the microphone. The results are given in below with test status and the estimated uncertainties. The "Pass" means the result of the test is inside the tolerances stated in the test specifications. The "-" means the result of test is outside these tolerances.

| Test: | Subtest: | Status: | Expanded Uncertainty (dB) | Coverage Factor |
|-------------------------|--|------------------------------|---------------------------|-----------------|
| Self-generated noise | A | Pass | 0.3 | |
| | C | Pass | 1.0 | 2.1 |
| | Lin | Pass | 2.0 | 2.2 |
| Linearity range for Leq | At reference range, Step 5 dB at 4 kHz | Pass | 0.3 | |
| | Reference SPL on all other ranges | Pass | 0.3 | |
| | 2 dB below upper limit of each range | Pass | 0.3 | |
| | 2 dB above lower limit of each range | Pass | 0.3 | |
| Linearity range for SPL | At reference range, Step 5 dB at 4 kHz | Pass | 0.3 | |
| | Frequency weightings | A | Pass | 0.3 |
| Time weightings | C | Pass | 0.3 | |
| | Lin | Pass | 0.3 | |
| | Single Burst Fast | Pass | 0.3 | |
| Peak response | Single Burst Slow | Pass | 0.3 | |
| | Single 100µs rectangular pulse | Pass | 0.3 | |
| R.M.S. accuracy | Crest factor of 3 | Pass | 0.3 | |
| | Time weighting I | Single burst 5 ms at 2000 Hz | Pass | 0.3 |
| Time averaging | Repeated at frequency of 100 Hz | Pass | 0.3 | |
| | 1 ms burst duty factor 1/10 ³ at 4kHz | Pass | 0.3 | |
| | 1 ms burst duty factor 1/10 ⁴ at 4kHz | Pass | 0.3 | |
| Pulse range | Single burst 10 ms at 4 kHz | Pass | 0.4 | |
| | Sound exposure level | Single burst 10 ms at 4 kHz | Pass | 0.4 |
| Overload indication | SPL | Pass | 0.3 | |
| | Leq | Pass | 0.4 | |

2, Acoustic tests

The complete sound level meter was calibrated on the reference range using a B&K 4226 acoustic calibrator with 1000Hz and SPL 94 dB. The sensitivity of the sound level meter was adjusted. The test result at 125 Hz and 8000 Hz are given in below with test status and the estimated uncertainties.

| Test: | Subtest | Status | Expanded Uncertainty (dB) | Coverage Factor |
|-------------------|------------------------|--------|---------------------------|-----------------|
| Acoustic response | Weighting A at 125 Hz | Pass | 0.3 | |
| | Weighting A at 8000 Hz | Pass | 0.5 | |

3, Response to associated sound calibrator

N/A

The expanded uncertainties have been calculated in accordance with the ISO Publication "Guide to the expression of uncertainty in measurement", and gives an interval estimated to have a level of confidence of 95%. A coverage factor of 2 is assumed unless explicitly stated.

Calibrated by:

Date: 29-May-2014

Fung Chi Yip

- End -

Checked by:

Date: 30-May-2014

Lam Tze Wai

The standard(s) and equipment used in the calibration are traceable to national or international recognised standards and are calibrated on a schedule to maintain the required accuracy level.



Test Data for Sound Level Meter

Sound level meter type: 2236 Serial No. 2100736 Date 29-May-2014
Microphone type: 4188 Serial No. 2157055
Report: 14CA0529 01-01

| | | | | | | |
|---------|------|------|------|-----|-----|------|
| 7943.0 | 94.0 | 91.0 | 90.9 | 1.5 | 3.0 | -0.1 |
| 12590.0 | 94.0 | 87.8 | 87.7 | 3.0 | 6.0 | -0.1 |

Frequency weighting Lin:

| Frequency Hz | Ref. level dB | Expected level dB | Actual level dB | Tolerance(dB) | | Deviation dB |
|-----------------|------------------|----------------------|--------------------|---------------|-----|-----------------|
| | | | | + | - | |
| 1000.0 | 94.0 | 94.0 | 94.0 | 0.0 | 0.0 | 0.0 |
| 31.6 | 94.0 | 94.0 | 94.0 | 1.5 | 1.5 | 0.0 |
| 63.1 | 94.0 | 94.0 | 93.9 | 1.5 | 1.5 | -0.1 |
| 125.9 | 94.0 | 94.0 | 94.0 | 1.0 | 1.0 | 0.0 |
| 251.2 | 94.0 | 94.0 | 94.0 | 1.0 | 1.0 | 0.0 |
| 501.2 | 94.0 | 94.0 | 94.0 | 1.0 | 1.0 | 0.0 |
| 1995.0 | 94.0 | 94.0 | 94.0 | 1.0 | 1.0 | 0.0 |
| 3981.0 | 94.0 | 94.0 | 94.0 | 1.0 | 1.0 | 0.0 |
| 7943.0 | 94.0 | 94.0 | 94.1 | 1.5 | 3.0 | 0.1 |
| 12590.0 | 94.0 | 94.0 | 94.1 | 3.0 | 6.0 | 0.1 |

TIME WEIGHTING FAST TEST

Time weighting F is tested on the reference range with a single sinusoidal burst of duration 200 ms at a frequency 2000 Hz and an amplitude which produces an indication 4 dB below the upper limit of the primary indicator range when the signal is continuous. (Weight A, Maximum hold)

| Ref. level dB | Expected level dB | Actual level dB | Tolerance(dB) | | Deviation dB |
|------------------|----------------------|--------------------|---------------|-----|-----------------|
| | | | + | - | |
| 109.0 | 108.0 | 108.0 | 1.0 | 1.0 | 0.0 |

TIME WEIGHTING SLOW TEST

Time weighting S is tested on the reference range with a single sinusoidal burst of duration 500 ms at a frequency 2000 Hz and an amplitude which produces an indication 4 dB below the upper limit of the primary indicator range when the signal is continuous. (Weight A, Maximum hold)

| Ref. level dB | Expected level dB | Actual level dB | Tolerance(dB) | | Deviation dB |
|------------------|----------------------|--------------------|---------------|-----|-----------------|
| | | | + | - | |
| 109.0 | 104.9 | 105.2 | 1.0 | 1.0 | 0.3 |

PEAK RESPONSE TEST

The onset time of the peak detector is tested on the reference range by comparing the response to a 100 us rectangular test pulse with the response to a 10 ms reference pulse of the same amplitude. The amplitude of the 10 ms reference pulse is such as to produce an indication 1 dB below the upper limit of the primary indicator range.

Positive polarities: (Weighting C, set the generator signal to single, Lcpmax)

| Ref. level dB | Response to 10 ms dB | Response to 100 us dB | Tolerance +/- dB | Deviation dB |
|------------------|-------------------------|--------------------------|---------------------|-----------------|
| 112.0 | 112.0 | 111.7 | 2.0 | -0.3 |

Negative polarities:

| Ref. level dB | Response to 10 ms dB | Response to 100 us dB | Tolerance +/- dB | Deviation dB |
|------------------|-------------------------|--------------------------|---------------------|-----------------|
| 112.0 | 112.0 | 111.7 | 2.0 | -0.3 |

RMS ACCURACY TEST

The RMS detector accuracy is tested on the reference range for a crest factor of 3.

Test frequency: 2000 Hz
Amplitude: 2 dB below the upper limit of the primary indicator range.



Test Data for Sound Level Meter

Page 4 of 5

Sound level meter type: 2236 Serial No. 2100736 Date 29-May-2014
Microphone type: 4188 Serial No. 2157055
Report: 14CA0529 01-01

| | | | | | |
|-----------------------------|---|----------------|-------------------|-----------|-----------|
| Burst repetition frequency: | 40 Hz | | | | |
| Tone burst signal: | 11 cycles of a sine wave of frequency 2000 Hz. (Set to INT) | | | | |
| | Ref. Level | Expected level | Tone burst signal | Tolerance | Deviation |
| Time weighting | dB | dB | indication(dB) | +/- dB | dB |
| Slow | 111.0+6.6 | 111.0 | 110.8 | 0.5 | -0.2 |

TIME WEIGHTING IMPULSE TEST

Time weighting | is tested on the reference range (Set the SLM to LAImax)

Test frequency: 2000 Hz

Amplitude: The upper limit of the primary indicator range.

Single sinusoidal burst of duration 5 ms:

| | | | | |
|------------|-------------------------|-------------|-----------|-----------|
| Ref. Level | Single burst indication | | Tolerance | Deviation |
| dB | Expected (dB) | Actual (dB) | +/- dB | dB |
| 113.0 | 104.2 | 104.2 | 2.0 | 0.0 |

Repeated at 100 Hz

| | | | | |
|------------|---------------------------|-------------|-----------|-----------|
| Ref. Level | Repeated burst indication | | Tolerance | Deviation |
| dB | Expected (dB) | Actual (dB) | +/- dB | dB |
| 113.0 | 110.3 | 110.3 | 1.0 | 0.0 |

TIME AVERAGING TEST

This test compares the SLM reading for continuous sine signals with readings obtained from a sine tone burst sequence having the same RMS level. The test level is 30 dB below the upper limit of the linearity range and repeated for Type 1 SLM with 40 dB below the upper limit of the linearity.

Frequency of tone burst: 4000 Hz

Duration of tone burst: 1 ms

| | | | | | | |
|-----------------|---------------------|--------------|------------|-----------|-----------|--------------|
| Repetition Time | Level of tone burst | Expected Leq | Actual Leq | Tolerance | Deviation | Remarks |
| msec | dB | dB | dB | +/- dB | dB | |
| 1000 | 83.0 | 83.0 | 82.7 | 1.0 | -0.3 | 60s integ. |
| 10000 | 73.0 | 73.0 | 72.7 | 1.0 | -0.3 | 6min. integ. |

PULSE RANGE AND SOUND EXPOSURE LEVEL TEST

The test tone burst signal is superimposed on a baseline signal corresponding to the lower limit of reference range

Test frequency: 4000 Hz

Integration time: 10 sec

The integrating sound level meter set to Leq:

| | | | | | |
|----------|------------------------------|----------|--------|-----------|-----------|
| Duration | Rms level of tone burst (dB) | Expected | Actual | Tolerance | Deviation |
| msec | dB | dB | dB | +/- dB | dB |
| 10 | 116.0 | 86.0 | 85.8 | 1.7 | -0.2 |

The integrating sound level meter set to SEL:

| | | | | | |
|----------|------------------------------|----------|--------|-----------|-----------|
| Duration | Rms level of tone burst (dB) | Expected | Actual | Tolerance | Deviation |
| msec | dB | dB | dB | +/- dB | dB |
| 10.0 | 116.0 | 96.0 | 95.9 | 1.7 | -0.1 |

OVERLOAD INDICATION TEST

For SLM capable of operating in a non-integrating mode.

Test frequency: 2000 Hz



Test Data for Sound Level Meter

Sound level meter type: 2236 Serial No. 2100736 Date 29-May-2014
Microphone type: 4188 Serial No. 2157055
Report: 14CA0529 01-01

Amplitude: 2 dB below the upper limit of the primary indicator range.
Burst repetition frequency: 40 Hz
Tone burst signal: 11 cycles of a sine wave of frequency 2000 Hz.

| Level | Level reduced by | Further reduced | Difference | Tolerance | Deviation |
|------------------|------------------|-----------------|------------|-----------|-----------|
| at overload (dB) | 1 dB | 3 dB | dB | dB | dB |
| 126.7 | 125.7 | 122.7 | 3.0 | 1.0 | 0.0 |

For integrating SLM, with the instrument indicating Leq.

For integrating SLM, with the instrument indicating Leq and set to the reference range. The test signal as following:
The test tone burst signal is superimposed on a baseline signal corresponding to the lower limit of reference range

Test frequency: 4000 Hz
Integration time: 10 sec
Single burst duration: 1 msec

| Rms level | Level reduced by | Expected level | Actual level | Tolerance | Deviation |
|------------------|------------------|----------------|--------------|-----------|-----------|
| at overload (dB) | 1 dB | dB | dB | dB | dB |
| 130.6 | 129.6 | 89.6 | 89.4 | 2.2 | -0.2 |

ACOUSTIC TEST

The acoustic test of the complete SLM is tested at the frequency 125 Hz and 8000 Hz using a B&K type 4226 Multifunction Acoustic Calibrator. The test is performed in A weighting.

| Frequency Hz | Expected level dB | Actual level Measured (dB) | Tolerance (dB) | | Deviation dB |
|-----------------|----------------------|-------------------------------|----------------|-----|-----------------|
| | | | + | - | |
| 1000 | 94.0 | 94.0 | 0.0 | 0.0 | 0.0 |
| 125 | 77.9 | 78.2 | 1.0 | 1.0 | 0.3 |
| 8000 | 92.9 | 92.8 | 1.5 | 3.0 | -0.1 |

-----END-----



CERTIFICATE OF CALIBRATION

Certificate No.: 14CA0529 01-02

Page: 1 of 2

Item tested

Description: Acoustical Calibrator (Class 1)
Manufacturer: Rion Co., Ltd.
Type/Model No.: NC-73
Serial/Equipment No.: 10465798
Adaptors used: -

Item submitted by

Customer: Lam Geotechnics Limited
Address of Customer: -
Request No.: -
Date of receipt: 29-May-2014

Date of test: 30-May-2014

Reference equipment used in the calibration

| Description: | Model: | Serial No. | Expiry Date: | Traceable to: |
|-------------------------|----------|------------|--------------|---------------|
| Lab standard microphone | B&K 4180 | 2412857 | 13-May-2015 | SCL |
| Preamplifier | B&K 2673 | 2239857 | 10-Apr-2015 | CEPREI |
| Measuring amplifier | B&K 2610 | 2346941 | 08-Apr-2015 | CEPREI |
| Signal generator | DS 360 | 61227 | 09-Apr-2015 | CEPREI |
| Digital multi-meter | 34401A | US36087050 | 17-Dec-2014 | CEPREI |
| Audio analyzer | 8903B | GB41300350 | 07-Apr-2015 | CEPREI |
| Universal counter | 53132A | MY40003662 | 11-Apr-2015 | CEPREI |

Ambient conditions

Temperature: 22 ± 1 °C
Relative humidity: 60 ± 10 %
Air pressure: 1000 ± 10 hPa

Test specifications

- The Sound Calibrator has been calibrated in accordance with the requirements as specified in IEC 60942 1997 Annex B and the lab calibration procedure SMTP004-CA-156.
- The calibrator was tested with its axis vertical facing downwards at the specific frequency using insert voltage technique.
- The results are rounded to the nearest 0.01 dB and 0.1 Hz and have not been corrected for variations from a reference pressure of 1013.25 hectoPascals as the maker's information indicates that the instrument is insensitive to pressure changes.

Test results

Details of the performed measurements are presented on **page 2** of this certificate.

Approved Signatory:


Huang Jian Min/Feng Jun Qi

Date: 30-May-2014

Company Chop:



Comments: The results reported in this certificate refer to the condition of the instrument on the date of calibration and carry no implication regarding the long-term stability of the instrument.



CERTIFICATE OF CALIBRATION

(Continuation Page)

Certificate No.: 14CA0529 01-02

Page: 2 of 2

1, Measured Sound Pressure Level

The output Sound Pressure Level in the calibrator head was measured at the setting and frequency shown using a calibrated laboratory standard microphone and insert voltage technique. The results are given in below with the estimated uncertainties.

| Frequency Shown Hz | Output Sound Pressure Level Setting dB | Measured Output Sound Pressure Level dB | (Output level in dB re 20 μ Pa) |
|-----------------------|---|--|--------------------------------------|
| | | | Estimated Expanded Uncertainty dB |
| 1000 | 94.00 | 94.57 | 0.10 |

2, Sound Pressure Level Stability - Short Term Fluctuations

The Short Term Fluctuations was determined by measuring the maximum and minimum of the fast weighted DC output of the B&K 2610 measuring amplifier over a 20 second time interval as required in the standard. The Short Term Fluctuation was found to be:

At 1000 Hz **STF = 0.001 dB**
Estimated expanded uncertainty 0.005 dB

3, Actual Output Frequency

The determination of actual output frequency was made using a B&K 4180 microphone together with a B&K 2673 preamplifier connected to a B&K 2610 measuring amplifier. The AC output of the B&K 2610 was taken to an universal counter which was used to determine the frequency averaged over 20 second of operation as required by the standard. The actual output frequency at 1 KHz was:

At 1000 Hz **Actual Frequency = 965.6 Hz**
Estimated expanded uncertainty 0.1 Hz Coverage factor k = 2.2

4, Total Noise and Distortion

For the Total Noise and Distortion measurement, the unfiltered AC output of the B&K 2610 measuring amplifier was connected to an Agilent Type 8903 B distortion analyser. The TND result at 1 KHz was:

At 1000 Hz **TND = 0.9 %**
Estimated expanded uncertainty 0.7 %

The expanded uncertainties have been calculated in accordance with the ISO Publication "Guide to the expression of uncertainty in measurement", and gives an interval estimated to have a level of confidence of 95%. A coverage factor of 2 is assumed unless explicitly stated.

Calibrated by:

Date: 30-May-2014

Fung Chi Yip

- End -

Checked by:

Date: 30-May-2014

Lam Tze Wai

The standard(s) and equipment used in the calibration are traceable to national or international recognised standards and are calibrated on a schedule to maintain the required accuracy level.

**REPORT OF EQUIPMENT PERFORMANCE CHECK/ CALIBRATION****WORK ORDER: HK1410093****DATE OF ISSUE: 28th May, 2014****CLIENT: LAM GEOTECHNICS LIMITED**

| | |
|----------------------------------|-----------------|
| Equipment Type: | Turbidimeter |
| Brand Name: | Xin Rui |
| Model No.: | WGZ-3B |
| Serial No.: | 1203010 |
| Equipment No.: | -- |
| Date of Calibration: | 28 May, 2014 |
| Date of next Calibration: | 28 August, 2014 |

Parameters:**Turbidity**Method Ref: APHA 22nd ed. 2130B

| Expected Reading (NTU) | Displayed Reading (NTU) | Tolerance (%) |
|------------------------|-----------------------------|---------------|
| 0 | 0.02 | --- |
| 4 | 4.16 | +4.0 |
| 10 | 9.80 | -2.0 |
| 40 | 38.5 | -3.75 |
| 100 | 104 | +4.0 |
| 400 | 420 | +5.0 |
| 1000 | 970 | -3.0 |
| | Tolerance Limit ($\pm\%$) | 10.0 |

Remark: "Displayed Reading" presents the figures shown on item under calibration/checking regardless of equipment precision or significant figures.

This report may not be reproduced except with prior written approval from Pilot Testing Limited.

Address: Room 1503, 15/F, Wayson Commercial House, 68-70 Lockhart Road, Wanchai, Hong Kong
Phone +852 2527 6691 | Email info@pilot-testing.com

**REPORT OF EQUIPMENT PERFORMANCE CHECK/ CALIBRATION****Information supplied by customer:**

CONTACT: DEREK LO **WORK ORDER:** HK1410074
CLIENT: LAM GEOTECHNICS LIMITED
DATE RECEIVED: 30/04/2014
DATE OF ISSUE: 04/05/2014
ADDRESS: 11/F, CENTRE POINT, 181-185, GLOUCESTER ROAD,
WANCHAI, HONG KONG

PROJECT: ---**METHOD OF PERFORMANCE CHECK/ CALIBRATION:**

Ref: APHA22nd ed 2130B

COMMENTS

It is certified that the item under performance check/calibration has been calibrated/checked by corresponding calibrated equipment in the laboratory. Maximum Tolerance and calibration frequency stated in the report, unless otherwise stated, the internal acceptance criteria of Pilot Testing Limited will be followed.

| | |
|-----------------------------|--------------|
| Scope of Test: | Turbidity |
| Equipment Type: | Turbidimeter |
| Brand Name: | Xin Rui |
| Model No.: | WGZ-3B |
| Serial No.: | 1203016 |
| Equipment No.: | -- |
| Date of Calibration: | 04 May, 2014 |

Remarks:

This is the Final Report. Results apply to sample(s) as submitted. All pages of this report have been checked and approved for release.

Mr. Peter Lee
Director

This report may not be reproduced except with prior written approval from Pilot Testing Limited.

Address: Room 1503, 15/F, Wayson Commercial House, 68-70 Lockhart Road, Wanchai, Hong Kong
Phone +852 2527 6691 | Email info@pilot-testing.com

**REPORT OF EQUIPMENT PERFORMANCE CHECK/ CALIBRATION****WORK ORDER: HK1410074****DATE OF ISSUE: 04th May, 2014****CLIENT: LAM GEOTECHNICS LIMITED**

| | |
|----------------------------------|-----------------|
| Equipment Type: | Turbidimeter |
| Brand Name: | Xin Rui |
| Model No.: | WGZ-3B |
| Serial No.: | 1203016 |
| Equipment No.: | -- |
| Date of Calibration: | 04 May, 2014 |
| Date of next Calibration: | 04 August, 2014 |

Parameters:**Turbidity**Method Ref: APHA 22nd ed. 2130B

| Expected Reading (NTU) | Displayed Reading (NTU) | Tolerance (%) |
|------------------------|-----------------------------|---------------|
| 0 | 0.02 | --- |
| 4 | 3.90 | -2.5 |
| 10 | 10.1 | +1.0 |
| 40 | 41.0 | +2.5 |
| 100 | 96.0 | -4.0 |
| 400 | 414 | +3.5 |
| 1000 | 970 | -3.0 |
| | Tolerance Limit ($\pm\%$) | 10.0 |

Remark: "Displayed Reading" presents the figures shown on item under calibration/checking regardless of equipment precision or significant figures.

This report may not be reproduced except with prior written approval from Pilot Testing Limited.

Address: Room 1503, 15/F, Wayson Commercial House, 68-70 Lockhart Road, Wanchai, Hong Kong
Phone +852 2527 6691 | Email info@pilot-testing.com

REPORT OF EQUIPMENT PERFORMANCE CHECK/ CALIBRATION**Information supplied by customer:****CONTACT: DEREK LO WORK ORDER: HK1410073****CLIENT: LAM GEOTECHNICS LIMITED****DATE RECEIVED: 30/04/2014****DATE OF ISSUE: 04/05/2014****ADDRESS: 11/E, CENTRE POINT, 181-185, GLOUCESTER ROAD,
WANCHAI, HONG KONG****PROJECT: ---****METHOD OF PERFORMANCE CHECK/ CALIBRATION:**

Ref: APHA22nd ed 2130B

COMMENTS

It is certified that the item under performance check/calibration has been calibrated/checked by corresponding calibrated equipment in the laboratory. Maximum Tolerance and calibration frequency stated in the report, unless otherwise stated, the internal acceptance criteria of Pilot Testing Limited will be followed.

| | |
|-----------------------------|--------------|
| Scope of Test: | Turbidity |
| Equipment Type: | Turbidimeter |
| Brand Name: | Xin Rui |
| Model No.: | WGZ-3B |
| Serial No.: | 1203025 |
| Equipment No.: | -- |
| Date of Calibration: | 04 May, 2014 |

Remarks:

This is the Final Report. Results apply to sample(s) as submitted. All pages of this report have been checked and approved for release.



Mr. Peter Lee

Director

This report may not be reproduced except with prior written approval from Pilot Testing Limited.

Address: Room 1503, 15/F, Wayson Commercial House, 68-70 Lockhart Road, Wanchai, Hong Kong
Phone +852 2527 6691 | Email info@pilot-testing.com

**REPORT OF EQUIPMENT PERFORMANCE CHECK/ CALIBRATION****WORK ORDER: HK1410073****DATE OF ISSUE: 04th May, 2014****CLIENT: LAM GEOTECHNICS LIMITED**

| | |
|----------------------------------|-----------------|
| Equipment Type: | Turbidimeter |
| Brand Name: | Xin Rui |
| Model No.: | WGZ-3B |
| Serial No.: | 1203025 |
| Equipment No.: | -- |
| Date of Calibration: | 04 May, 2014 |
| Date of next Calibration: | 04 August, 2014 |

Parameters:**Turbidity**Method Ref: APHA 22nd ed. 2130B

| Expected Reading (NTU) | Displayed Reading (NTU) | Tolerance (%) |
|------------------------|-----------------------------|---------------|
| 0 | 0.02 | --- |
| 4 | 3.86 | -3.5 |
| 10 | 10.3 | +3.0 |
| 40 | 42.0 | +5.0 |
| 100 | 97.0 | -3.0 |
| 400 | 406 | +1.5 |
| 1000 | 975 | -2.5 |
| | Tolerance Limit ($\pm\%$) | 10.0 |

Remark: "Displayed Reading" presents the figures shown on item under calibration/checking regardless of equipment precision or significant figures.

This report may not be reproduced except with prior written approval from Pilot Testing Limited.

Address: Room 1503, 15/F, Wayson Commercial House, 68-70 Lockhart Road, Wanchai, Hong Kong
Phone +852 2527 6691 | Email info@pilot-testing.com



ALS Technichem (HK) Pty Ltd
11/F, Chung Shun Knitting Centre
1-3 Wing Yip Street
Kwai Chung, N.T., Hong Kong
T: +852 2610 1044
F: +852 2610 2021
www.alsglobal.com

REPORT OF EQUIPMENT PERFORMANCE CHECK/CALIBRATION

CONTACT: MS PAULINE WONG
CLIENT: LAM ENVIRONMENTAL SERVICES LTD
ADDRESS: 11/F., CENTRE POINT,
181-185 GLOUCESTER ROAD,
WAN CHAI, HONG KONG
PROJECT: --

WORK ORDER: HK1412271
LABORATORY: HONG KONG
DATE RECEIVED: 22/04/2014
DATE OF ISSUE: 02/05/2014

COMMENTS

It is certified that the item under calibration/checking has been calibrated/checked by corresponding calibrated equipment in the laboratory.

Maximum Tolerance and calibration frequency stated in the report, unless otherwise stated, the internal acceptance criteria of ALS will be followed.

Scope of Test: Dissolved Oxygen, pH, Salinity and Temperature
Description: Multimeter
Brand Name: YSI
Model No.: PROFESSIONAL PLUS
Serial No.: 11F100597
Equipment No.: --
Date of Calibration: 29 April, 2014

NOTES

This is the Final Report and supersedes any preliminary report with this batch number.

Results apply to sample(s) as submitted. All pages of this report have been checked and approved for release.


Mr. Fung Lim Chee, Richard
General Manager -
Greater China & Hong Kong



TISCH ENVIRONMENTAL, INC.
 145 SOUTH MIAMI AVE.
 VILLAGE OF CLEVELAND, OH 45002
 513.467.9000
 877.263.7610 TOLL FREE
 513.467.9009 FAX
 WWW.TISCH-ENV.COM

AIR POLLUTION MONITORING EQUIPMENT
 ORIFICE TRANSFER STANDARD CERTIFICATION WORKSHEET TE-5025A

Date - Jul 15, 2013 Roots-meter S/N 0438320 Ta (K) - 300
 Operator Tisch Orifice I.D. - 0005 Pa (mm) - 759.46

| PLATE OR Run # | VOLUME START (m3) | VOLUME STOP (m3) | DIFF VOLUME (m3) | DIFF TIME (min) | METER | ORFICE |
|----------------|-------------------|------------------|------------------|-----------------|--------------|----------------|
| | | | | | DIFF Hg (mm) | DIFF H2O (in.) |
| 1 | NA | NA | 1.00 | 1.3910 | 3.2 | 2.00 |
| 2 | NA | NA | 1.00 | 0.9830 | 6.4 | 4.00 |
| 3 | NA | NA | 1.00 | 0.8800 | 7.9 | 5.00 |
| 4 | NA | NA | 1.00 | 0.8380 | 8.8 | 5.50 |
| 5 | NA | NA | 1.00 | 0.6930 | 12.7 | 8.00 |

DATA TABULATION

| Vstd | (x axis) Qstd | (y axis) | Va | (x axis) Qa | (y axis) |
|--|---------------|----------|-------------------------------|-------------|----------|
| 0.9884 | 0.7106 | 1.4090 | 0.9958 | 0.7159 | 0.8888 |
| 0.9843 | 1.0013 | 1.9926 | 0.9916 | 1.0087 | 1.2570 |
| 0.9822 | 1.1161 | 2.2278 | 0.9895 | 1.1244 | 1.4054 |
| 0.9811 | 1.1708 | 2.3365 | 0.9884 | 1.1795 | 1.4740 |
| 0.9760 | 1.4084 | 2.8180 | 0.9832 | 1.4188 | 1.7777 |
| Qstd slope (m) = 2.01968 | | | Qa slope (m) = 1.26469 | | |
| intercept (b) = -0.02746 | | | intercept (b) = -0.01732 | | |
| coefficient (r) = 0.99999 | | | coefficient (r) = 0.99999 | | |
| y axis = $\sqrt{H_2O(Pa/760)(298/Ta)}$ | | | y axis = $\sqrt{H_2O(Ta/Pa)}$ | | |

CALCULATIONS

$$Vstd = \text{Diff. Vol} [(Pa - \text{Diff. Hg}) / 760] (298 / Ta)$$

$$Qstd = Vstd / \text{Time}$$

$$Va = \text{Diff Vol} [(Pa - \text{Diff Hg}) / Pa]$$

$$Qa = Va / \text{Time}$$

For subsequent flow rate calculations:

$$Qstd = 1/m \{ [\sqrt{H_2O(Pa/760)(298/Ta)}] - b \}$$

$$Qa = 1/m \{ [\sqrt{H_2O(Ta/Pa)}] - b \}$$

REPORT OF EQUIPMENT PERFORMANCE CHECK/CALIBRATION

Work Order: HK1412271
 Date of Issue: 02/05/2014
 Client: LAM ENVIRONMENTAL SERVICES LTD



Description: Multimeter
 Brand Name: YSI
 Model No.: PROFESSIONAL PLUS
 Serial No.: 11F100597
 Equipment No.: --
 Date of Calibration: 29 April, 2014

Date of next Calibration: 29 July, 2014

Parameters:

Dissolved Oxygen Method Ref: APHA (21st edition), 4500O: G

| Expected Reading (mg/L) | Displayed Reading (mg/L) | Tolerance (mg/L) |
|-------------------------|--------------------------|------------------|
| 3.94 | 3.88 | -0.06 |
| 6.10 | 5.90 | -0.20 |
| 7.98 | 7.89 | -0.09 |
| Tolerance Limit (mg/L) | | ±0.20 |

pH Value

Method Ref: APHA (21st edition), 4500H:B

| Expected Reading (pH Unit) | Displayed Reading (pH Unit) | Tolerance (pH unit) |
|----------------------------|-----------------------------|---------------------|
| 4.0 | 4.16 | +0.16 |
| 7.0 | 7.13 | +0.13 |
| 10.0 | 10.06 | +0.06 |
| Tolerance Limit (pH Unit) | | ±0.20 |

Salinity

Method Ref: APHA (21st edition), 2520B

| Expected Reading (g/L) | Displayed Reading (g/L) | Tolerance (%) |
|------------------------|-------------------------|---------------|
| 0 | 0.00 | -- |
| 10 | 9.12 | -8.8 |
| 20 | 18.80 | -6.0 |
| 30 | 27.70 | -7.7 |
| Tolerance Limit (%) | | ±10.0 |

Temperature

Method Ref: Section 6 of International Accreditation New Zealand Technical

Guide No. 3 Second edition March 2008: Working Thermometer Calibration Procedure.

| Reading of Ref. thermometer (°C) | Displayed Reading (°C) | Tolerance (°C) |
|----------------------------------|------------------------|----------------|
| 10.5 | 10.2 | -0.3 |
| 25.5 | 25.3 | -0.2 |
| 37.5 | 37.5 | 0.0 |
| Tolerance Limit (°C) | | ±2.0 |

Remark: "Displayed Reading" presents the figures shown on item under calibration / checking regardless of equipment precision or significant figures.


 Mr. Fung Lim Chee, Richard
 General Manager -
 Greater China & Hong Kong



ALS Technichem (HK) Pty Ltd
11/F, Chung Shun Knitting Centre
1-3 Wing Yip Street
Kwai Chung, N.T., Hong Kong
T: +852 2610 1044
F: +852 2610 2021
www.alsglobal.com

REPORT OF EQUIPMENT PERFORMANCE CHECK/CALIBRATION

CONTACT: MS PAULINE WONG
CLIENT: LAM GEOTECHNICS LIMITED
ADDRESS: 11/F., CENTRE POINT,
181-185 GLOUCESTER ROAD,
WAN CHAI, HONG KONG

WORK ORDER: HK1411576
LABORATORY: HONG KONG
DATE RECEIVED: 14/04/2014
DATE OF ISSUE: 17/04/2014

PROJECT: --

COMMENTS

It is certified that the item under calibration/checking has been calibrated/checked by corresponding calibrated equipment in the laboratory.

Maximum Tolerance and calibration frequency stated in the report, unless otherwise stated, the internal acceptance criteria of ALS will be followed.

Scope of Test: pH, Temperature, Salinity and Dissolved Oxygen
Description: Multimeter
Brand Name: YSI
Model No.: Professional Plus
Serial No.: 11F100420
Equipment No.: --
Date of Calibration: 17 April, 2014

NOTES

This is the Final Report and supersedes any preliminary report with this batch number.

Results apply to sample(s) as submitted. All pages of this report have been checked and approved for release.


Mr. Fung Lim Chee, Richard
General Manager
Greater China & Hong Kong

REPORT OF EQUIPMENT PERFORMANCE CHECK/CALIBRATION

Work Order: HK1411576
Date of Issue: 17/04/2014
Client: LAM GEOTECHNICS LIMITED



Description: Multimeter
Brand Name: YSI
Model No.: Professional Plus
Serial No.: 11F100420
Equipment No.: --
Date of Calibration: 17 April, 2014

Date of next Calibration: 17 July, 2014

Parameters:

Dissolved Oxygen Method Ref: APHA (21st edition), 4500O: G

| Expected Reading (mg/L) | Displayed Reading (mg/L) | Tolerance (mg/L) |
|-------------------------|--------------------------|------------------|
| 3.86 | 3.79 | -0.07 |
| 5.65 | 5.76 | +0.11 |
| 8.02 | 8.12 | +0.10 |
| Tolerance Limit (mg/L) | | ±0.20 |

pH Value

Method Ref: APHA (21st edition), 4500H:B

| Expected Reading (pH Unit) | Displayed Reading (pH Unit) | Tolerance (pH unit) |
|----------------------------|-----------------------------|---------------------|
| 4.0 | 3.97 | -0.03 |
| 7.0 | 6.92 | -0.08 |
| 10.0 | 9.97 | -0.03 |
| Tolerance Limit (pH Unit) | | ±0.20 |

Salinity

Method Ref: APHA (21st edition), 2520B

| Expected Reading (g/L) | Displayed Reading (g/L) | Tolerance (%) |
|------------------------|-------------------------|---------------|
| 0 | 0.00 | -- |
| 10 | 9.57 | -4.3 |
| 20 | 18.85 | -5.7 |
| 30 | 30.14 | +0.5 |
| Tolerance Limit (%) | | ±10.0 |

Temperature

Method Ref: Section 6 of International Accreditation New Zealand Technical

Guide No. 3 Second edition March 2008: Working Thermometer Calibration Procedure.

| Reading of Ref. thermometer (°C) | Displayed Reading (°C) | Tolerance (°C) |
|----------------------------------|------------------------|----------------|
| 9.5 | 9.9 | +0.4 |
| 22.0 | 22.1 | +0.1 |
| 39.0 | 39.3 | +0.3 |
| Tolerance Limit (°C) | | ±2.0 |

Remark: "Displayed Reading" presents the figures shown on item under calibration / checking regardless of equipment precision or significant figures.

Mr. Fung Lim Chee, Richard
 General Manager -
 Greater China & Hong Kong



ALS Technichem (HK) Pty Ltd
11/F, Chung Shun Knitting Centre
1-3 Wing Yip Street
Kwai Chung, N.T., Hong Kong
T: +852 2610 1044
F: +852 2610 2021
www.alsglobal.com

REPORT OF EQUIPMENT PERFORMANCE CHECK/CALIBRATION

CONTACT: MS PAULINE WONG
CLIENT: LAM GEOTECHNICS LIMITED
ADDRESS: 11/F., CENTRE POINT,
181-185 GLOUCESTER ROAD,
WAN CHAI, HONG KONG

WORK ORDER: HK1418648
LABORATORY: HONG KONG
DATE RECEIVED: 13/06/2014
DATE OF ISSUE: 24/06/2014

PROJECT: --

COMMENTS

The performance of the equipment stated in this report is checked with independent reference material and results compared against a calibrated secondary source.

The "Tolerance Limit" quoted is the acceptance criteria applicable for similar equipment used by the ALS Hong Kong laboratory or quoted from relevant international standards.

The "Next Calibration Date" is recommended according to best practice principals as practised by the ALS Hong Kong laboratory or quoted from relevant international standards.

Scope of Test: pH, Temperature, Salinity and Dissolved Oxygen
Description: Multimeter
Brand Name: YSI
Model No.: Professional Plus
Serial No.: 13A100242
Equipment No.: --
Date of Calibration: 19 June, 2014

NOTES

This is the Final Report and supersedes any preliminary report with this batch number.
Results apply to sample(s) as submitted. All pages of this report have been checked and approved for release.


Mr. Fung Lim Chee, Richard
General Manager -
Greater China & Hong Kong

REPORT OF EQUIPMENT PERFORMANCE CHECK/CALIBRATION

Work Order: HK1418648
Date of Issue: 24/06/2014
Client: LAM GEOTECHNICS LIMITED



Description: Multimeter
Brand Name: YSI
Model No.: Professional Plus
Serial No.: 13A100242
Equipment No.: --
Date of Calibration: 19 June, 2014

Date of next Calibration: 19 September, 2014

Parameters:

Dissolved Oxygen Method Ref: APHA (21st edition), 4500O: G

| Expected Reading (mg/L) | Displayed Reading (mg/L) | Tolerance (mg/L) |
|-------------------------|--------------------------|------------------|
| 4.38 | 4.52 | +0.14 |
| 6.42 | 6.46 | +0.04 |
| 7.95 | 7.87 | -0.08 |
| Tolerance Limit (mg/L) | | ±0.20 |

pH Value

Method Ref: APHA (21st edition), 4500H:B

| Expected Reading (pH Unit) | Displayed Reading (pH Unit) | Tolerance (pH unit) |
|----------------------------|-----------------------------|---------------------|
| 4.0 | 4.07 | +0.07 |
| 7.0 | 7.02 | +0.02 |
| 10.0 | 10.13 | +0.13 |
| Tolerance Limit (pH Unit) | | ±0.20 |

Salinity

Method Ref: APHA (21st edition), 2520B

| Expected Reading (g/L) | Displayed Reading (g/L) | Tolerance (%) |
|------------------------|-------------------------|---------------|
| 0 | 0.00 | -- |
| 10 | 9.40 | -6.0 |
| 20 | 18.81 | -6.0 |
| 30 | 28.28 | -5.7 |
| Tolerance Limit (%) | | ±10.0 |

Temperature

Method Ref: Section 6 of International Accreditation New Zealand Technical

Guide No. 3 Second edition March 2008: Working Thermometer Calibration Procedure.

| Reading of Ref. thermometer (°C) | Displayed Reading (°C) | Tolerance (°C) |
|----------------------------------|------------------------|----------------|
| 12.2 | 12.1 | -0.1 |
| 24.4 | 24.2 | -0.2 |
| 33.7 | 33.6 | -0.1 |
| Tolerance Limit (°C) | | ±2.0 |

Remark: "Displayed Reading" presents the figures shown on item under calibration / checking regardless of equipment precision or significant figures.

 Mr. Fung Lim Chee, Richard
 General Manager
 Greater China & Hong Kong



TISCH ENVIRONMENTAL, INC.
 145 SOUTH MIAMI AVE.
 VILLAGE OF CLEVELAND, OH 45002
 513.467.9000
 877.263.7610 TOLL FREE
 513.467.9009 FAX
 WWW.TISCH-ENV.COM

AIR POLLUTION MONITORING EQUIPMENT
 ORIFICE TRANSFER STANDARD CERTIFICATION WORKSHEET TE-5025A

Date - Jul 15, 2013 Roots-meter S/N 0438320 Ta (K) - 300
 Operator Tisch Orifice I.D. - 0005 Pa (mm) - 759.46

| PLATE OR Run # | VOLUME START (m3) | VOLUME STOP (m3) | DIFF VOLUME (m3) | DIFF TIME (min) | METER | ORFICE |
|----------------|-------------------|------------------|------------------|-----------------|--------------|----------------|
| | | | | | DIFF Hg (mm) | DIFF H2O (in.) |
| 1 | NA | NA | 1.00 | 1.3910 | 3.2 | 2.00 |
| 2 | NA | NA | 1.00 | 0.9830 | 6.4 | 4.00 |
| 3 | NA | NA | 1.00 | 0.8800 | 7.9 | 5.00 |
| 4 | NA | NA | 1.00 | 0.8380 | 8.8 | 5.50 |
| 5 | NA | NA | 1.00 | 0.6930 | 12.7 | 8.00 |

DATA TABULATION

| Vstd | (x axis) Qstd | (y axis) | Va | (x axis) Qa | (y axis) |
|--|---------------|----------|---|-------------|----------|
| 0.9884 | 0.7106 | 1.4090 | 0.9958 | 0.7159 | 0.8888 |
| 0.9843 | 1.0013 | 1.9926 | 0.9916 | 1.0087 | 1.2570 |
| 0.9822 | 1.1161 | 2.2278 | 0.9895 | 1.1244 | 1.4054 |
| 0.9811 | 1.1708 | 2.3365 | 0.9884 | 1.1795 | 1.4740 |
| 0.9760 | 1.4084 | 2.8180 | 0.9832 | 1.4188 | 1.7777 |
| Qstd slope (m) = 2.01968 | | | Qa slope (m) = 1.26469 | | |
| intercept (b) = -0.02746 | | | intercept (b) = -0.01732 | | |
| coefficient (r) = 0.99999 | | | coefficient (r) = 0.99999 | | |
| y axis = $\text{SQRT}[\text{H}_2\text{O}(\text{Pa}/760)(298/\text{Ta})]$ | | | y axis = $\text{SQRT}[\text{H}_2\text{O}(\text{Ta}/\text{Pa})]$ | | |

CALCULATIONS

$$\text{Vstd} = \text{Diff. Vol} [(\text{Pa} - \text{Diff. Hg}) / 760] (298 / \text{Ta})$$

$$\text{Qstd} = \text{Vstd} / \text{Time}$$

$$\text{Va} = \text{Diff Vol} [(\text{Pa} - \text{Diff Hg}) / \text{Pa}]$$

$$\text{Qa} = \text{Va} / \text{Time}$$

For subsequent flow rate calculations:

$$\text{Qstd} = 1/m \{ [\text{SQRT}(\text{H}_2\text{O}(\text{Pa}/760)(298/\text{Ta}))] - b \}$$

$$\text{Qa} = 1/m \{ [\text{SQRT}(\text{H}_2\text{O}(\text{Ta}/\text{Pa}))] - b \}$$



Lam Geotechnics Limited

Calibration Data for High Volume Sampler (TSP Sampler)

Location : CMA1b Calibration Date : 13-May-14
 Equipment no. : EL452 Calibration Due Date : 13-Jul-14

CALIBRATION OF CONTINUOUS FLOW RECORDER

| Ambient Condition | | | |
|-----------------------------|-----|--------|--------------------------|
| Temperature, T _a | 300 | Kelvin | Pressure, P _a |
| | | | 1007 mmHg |

| Orifice Transfer Standard Information | | | |
|---------------------------------------|-----------|--|----------|
| Equipment No. | EL086 | Slope, m _c | 2.01968 |
| | | Intercept, b _c | -0.02746 |
| Last Calibration Date | 15-Jul-13 | $\left(\frac{H \times P_a}{1013.3 \times 298 / T_a} \right)^{1/2}$ $= m_c \times Q_{std} + b_c$ | |
| Next Calibration Date | 15-Jul-14 | | |

| Calibration of TSP | | | | | | |
|--------------------|-------------------|--------|--------------|---|--------------------------------------|--|
| Calibration Point | Manometer Reading | | | Q _{std} (m ³ / min.) X-axis | Continuous Flow Recorder, W (CFM) | IC $(W(P_a/1013.3 \times 298/T_a)^{1/2}/35.31)$ Y-axis |
| | (up) | (down) | (difference) | | | |
| 1 | 6.2 | 6.2 | 12.4 | 1.7459 | 61 | 60.6070 |
| 2 | 5.1 | 5.1 | 10.2 | 1.5847 | 51 | 50.6715 |
| 3 | 4.1 | 4.1 | 8.2 | 1.4223 | 43 | 42.7230 |
| 4 | 2.5 | 2.5 | 5.0 | 1.1136 | 27 | 26.8261 |
| 5 | 1.4 | 1.4 | 2.8 | 0.8368 | 14 | 13.9098 |

By Linear Regression of Y on X

Slope, m = 50.9704 Intercept, b = -29.3862
 Correlation Coefficient* = 0.9991
 Calibration Accepted = Yes/No**

* if Correlation Coefficient < 0.990, check and recalibration again.

** Delete as appropriate.

Remarks : _____

Calibrated by : Felix Li Checked by : Derek Lo
 Date : 13-May-14 Date : 13-May-14



Lam Geotechnics Limited

Calibration Data for High Volume Sampler (TSP Sampler)

Location : CMA1b
 Equipment no. : EL452

Calibration Date : 2-Jul-14
 Calibration Due Date : 2-Sep-14

CALIBRATION OF CONTINUOUS FLOW RECORDER

| Ambient Condition | | | |
|-----------------------------|-----|--------|--------------------------|
| Temperature, T _a | 302 | Kelvin | Pressure, P _a |
| | | | 1009 mmHg |

| Orifice Transfer Standard Information | | | | | |
|---------------------------------------|-----------|--|---------|---------------------------|----------|
| Equipment No. | EL086 | Slope, m _c | 2.01968 | Intercept, b _c | -0.02746 |
| Last Calibration Date | 15-Jul-13 | $\left(H \times P_a / 1013.3 \times 298 / T_a \right)^{1/2}$ $= m_c \times Q_{std} + b_c$ | | | |
| Next Calibration Date | 15-Jul-14 | | | | |

| Calibration of TSP | | | | | | |
|--------------------|-------------------|--------|--------------|---|--------------------------------------|---|
| Calibration Point | Manometer Reading | | | Q _{std} (m ³ / min.) X-axis | Continuous Flow Recorder, W (CFM) | IC (W(P _a /1013.3x298/T _a) ^{1/2} /35.31) Y-axis |
| | (up) | (down) | (difference) | | | |
| 1 | 6.3 | 6.3 | 12.6 | 1.7557 | 60 | 59.4747 |
| 2 | 5.0 | 5.0 | 10.0 | 1.5656 | 49 | 48.5710 |
| 3 | 4.0 | 4.0 | 8.0 | 1.4018 | 40 | 39.6498 |
| 4 | 2.7 | 2.7 | 5.4 | 1.1541 | 26 | 25.7724 |
| 5 | 1.2 | 1.2 | 2.4 | 0.7739 | 12 | 11.8949 |

By Linear Regression of Y on X

Slope, m = 48.8251 Intercept, b = -27.8761
 Correlation Coefficient* = 0.9947
 Calibration Accepted = Yes/No**

* if Correlation Coefficient < 0.990, check and recalibration again.

** Delete as appropriate.

Remarks : _____

Calibrated by : Henry Lau
 Date : 2-Jul-14

Checked by : Pauline Wong
 Date : 2-Jul-14



Lam Geotechnics Limited

Calibration Data for High Volume Sampler (TSP Sampler)

Location : CMA2a Calibration Date : 13-May-14
 Equipment no. : EL449 Calibration Due Date : 13-Jul-14

CALIBRATION OF CONTINUOUS FLOW RECORDER

| Ambient Condition | | | |
|-----------------------------|-----|--------|--------------------------|
| Temperature, T _a | 300 | Kelvin | Pressure, P _a |
| | | | 1007 mmHg |

| Orifice Transfer Standard Information | | | |
|---------------------------------------|-----------|--|----------|
| Equipment No. | EL086 | Slope, m _c | 2.01968 |
| | | Intercept, b _c | -0.02746 |
| Last Calibration Date | 15-Jul-13 | $\left(H \times P_a / 1013.3 \times 298 / T_a \right)^{1/2}$ $= m_c \times Q_{std} + b_c$ | |
| Next Calibration Date | 15-Jul-14 | | |

| Calibration of TSP | | | | | | |
|--------------------|-------------------|--------|--------------|---|--------------------------------------|---|
| Calibration Point | Manometer Reading | | | Q _{std} (m ³ / min.) X-axis | Continuous Flow Recorder, W (CFM) | IC (W(P _a /1013.3x298/T _a) ^{1/2} /35.31) Y-axis |
| | (up) | (down) | (difference) | | | |
| 1 | 6.1 | 6.1 | 12.2 | 1.7319 | 61 | 60.6070 |
| 2 | 5.2 | 5.2 | 10.4 | 1.6000 | 53 | 52.6586 |
| 3 | 4.0 | 4.0 | 8.0 | 1.4050 | 43 | 42.7230 |
| 4 | 2.4 | 2.4 | 4.8 | 1.0914 | 26 | 25.8325 |
| 5 | 1.4 | 1.4 | 2.8 | 0.8368 | 14 | 13.9098 |

By Linear Regression of Y on X

Slope, m = 52.1379 Intercept, b = -30.3543
 Correlation Coefficient* = 0.9995
 Calibration Accepted = Yes/No**

* if Correlation Coefficient < 0.990, check and recalibration again.

** Delete as appropriate.

Remarks : _____

Calibrated by : Felix Li Checked by : Derek Lo
 Date : 13-May-14 Date : 13-May-14



Lam Geotechnics Limited

Calibration Data for High Volume Sampler (TSP Sampler)

Location : CMA2a
 Equipment no. : EL449

Calibration Date : 2-Jul-14
 Calibration Due Date : 2-Sep-14

CALIBRATION OF CONTINUOUS FLOW RECORDER

| Ambient Condition | | | |
|-----------------------------|-----|--------|--------------------------|
| Temperature, T _a | 302 | Kelvin | Pressure, P _a |
| | | | 1009 mmHg |

| Orifice Transfer Standard Information | | | |
|---------------------------------------|-----------|--|----------|
| Equipment No. | EL086 | Slope, m _c | 2.01968 |
| | | Intercept, b _c | -0.02746 |
| Last Calibration Date | 15-Jul-13 | $\left(H \times P_a / 1013.3 \times 298 / T_a \right)^{1/2}$ $= m_c \times Q_{std} + b_c$ | |
| Next Calibration Date | 15-Jul-14 | | |

| Calibration of TSP | | | | | | |
|--------------------|-------------------|--------|--------------|---|--------------------------------------|---|
| Calibration Point | Manometer Reading | | | Q _{std} (m ³ / min.) X-axis | Continuous Flow Recorder, W (CFM) | IC (W(P _a /1013.3x298/T _a) ^{1/2} /35.31) Y-axis |
| | (up) | (down) | (difference) | | | |
| 1 | 6.1 | 6.1 | 12.2 | 1.7279 | 60 | 59.4747 |
| 2 | 4.7 | 4.7 | 9.4 | 1.5183 | 54 | 53.5273 |
| 3 | 3.7 | 3.7 | 7.4 | 1.3487 | 48 | 47.5798 |
| 4 | 2.3 | 2.3 | 4.6 | 1.0662 | 40 | 39.6498 |
| 5 | 1.4 | 1.4 | 2.8 | 0.8349 | 30 | 29.7374 |

By Linear Regression of Y on X

Slope, m = 32.7993 Intercept, b = 3.3810
 Correlation Coefficient* = 0.9971
 Calibration Accepted = Yes/No**

* if Correlation Coefficient < 0.990, check and recalibration again.

** Delete as appropriate.

Remarks : _____

Calibrated by : Henry Lau
 Date : 2-Jul-14

Checked by : Pauline Wong
 Date : 2-Jul-14



Lam Geotechnics Limited

Calibration Data for High Volume Sampler (TSP Sampler)

Location : CMA3a
 Equipment no. : EL333

Calibration Date : 21-Jun-14
 Calibration Due Date : 21-Aug-14

CALIBRATION OF CONTINUOUS FLOW RECORDER

| Ambient Condition | | | |
|-----------------------------|-----|--------|--------------------------|
| Temperature, T _a | 301 | Kelvin | Pressure, P _a |
| | | | 1003 mmHg |

| Orifice Transfer Standard Information | | | |
|---------------------------------------|-----------|--|----------|
| Equipment No. | EL086 | Slope, m _c | 2.01968 |
| | | Intercept, b _c | -0.02746 |
| Last Calibration Date | 15-Jul-13 | $\left(H \times P_a / 1013.3 \times 298 / T_a \right)^{1/2}$ $= m_c \times Q_{std} + b_c$ | |
| Next Calibration Date | 15-Jul-14 | | |

| Calibration of TSP | | | | | | |
|--------------------|-------------------|--------|--------------|---|--------------------------------------|---|
| Calibration Point | Manometer Reading | | | Q _{std} (m ³ / min.) X-axis | Continuous Flow Recorder, W (CFM) | IC (W(P _a /1013.3x298/T _a) ^{1/2} /35.31) Y-axis |
| | (up) | (down) | (difference) | | | |
| 1 | 5.5 | 5.5 | 11.0 | 1.6392 | 61 | 60.3860 |
| 2 | 4.4 | 4.4 | 8.8 | 1.4676 | 52 | 51.4766 |
| 3 | 3.7 | 3.7 | 7.4 | 1.3469 | 43 | 42.5672 |
| 4 | 2.4 | 2.4 | 4.8 | 1.0874 | 26 | 25.7383 |
| 5 | 1.5 | 1.5 | 3.0 | 0.8626 | 14 | 13.8591 |

By Linear Regression of Y on X

Slope, m = 61.3367 Intercept, b = -39.7515
 Correlation Coefficient* = 0.9987
 Calibration Accepted = Yes/No**

* if Correlation Coefficient < 0.990, check and recalibration again.

** Delete as appropriate.

Remarks : _____

Calibrated by : Felix Li
 Date : 21-Jun-14

Checked by : Pauline Wong
 Date : 21-Jun-14



Lam Geotechnics Limited

Calibration Data for High Volume Sampler (TSP Sampler)

Location : CMA4a Calibration Date : 13-May-14
 Equipment no. : EL390 Calibration Due Date : 13-Jul-14

CALIBRATION OF CONTINUOUS FLOW RECORDER

| Ambient Condition | | | |
|-----------------------------|-----|--------|--------------------------|
| Temperature, T _a | 300 | Kelvin | Pressure, P _a |
| | | | 1007 mmHg |

| Orifice Transfer Standard Information | | | |
|---------------------------------------|-----------|---|----------|
| Equipment No. | EL086 | Slope, m _c | 2.01968 |
| | | Intercept, b _c | -0.02746 |
| Last Calibration Date | 15-Jul-13 | $(H \times P_a / 1013.3 \times 298 / T_a)^{1/2}$ $= m_c \times Q_{std} + b_c$ | |
| Next Calibration Date | 15-Jul-14 | | |

| Calibration of TSP | | | | | | |
|--------------------|-------------------|--------|--------------|---|--------------------------------------|---|
| Calibration Point | Manometer Reading | | | Q _{std} (m ³ / min.) X-axis | Continuous Flow Recorder, W (CFM) | IC (W(P _a /1013.3x298/T _a) ^{1/2} /35.31) Y-axis |
| | (up) | (down) | (difference) | | | |
| 1 | 6.0 | 6.0 | 12.0 | 1.7177 | 62 | 61.6006 |
| 2 | 5.1 | 5.1 | 10.2 | 1.5847 | 53 | 52.6586 |
| 3 | 4.0 | 4.0 | 8.0 | 1.4050 | 43 | 42.7230 |
| 4 | 2.6 | 2.6 | 5.2 | 1.1354 | 27 | 26.8261 |
| 5 | 1.5 | 1.5 | 3.0 | 0.8657 | 13 | 12.9163 |

By Linear Regression of Y on X

Slope, m = 56.9672 Intercept, b = -37.0880
 Correlation Coefficient* = 0.9993
 Calibration Accepted = Yes/No**

* if Correlation Coefficient < 0.990, check and recalibration again.

** Delete as appropriate.

Remarks : _____

Calibrated by : Felix Li Checked by : Derek Lo
 Date : 13-May-14 Date : 13-May-14



Lam Geotechnics Limited

Calibration Data for High Volume Sampler (TSP Sampler)

Location : CMA4a Calibration Date : 2-Jul-14
 Equipment no. : EL390 Calibration Due Date : 2-Sep-14

CALIBRATION OF CONTINUOUS FLOW RECORDER

| Ambient Condition | | | |
|-----------------------------|-----|--------|--------------------------|
| Temperature, T _a | 302 | Kelvin | Pressure, P _a |
| | | | 1009 mmHg |

| Orifice Transfer Standard Information | | | | | |
|---------------------------------------|-----------|---|---------|---------------------------|----------|
| Equipment No. | EL086 | Slope, m _c | 2.01968 | Intercept, b _c | -0.02746 |
| Last Calibration Date | 15-Jul-13 | $(H \times P_a / 1013.3 \times 298 / T_a)^{1/2}$ $= m_c \times Q_{std} + b_c$ | | | |
| Next Calibration Date | 15-Jul-14 | | | | |

| Calibration of TSP | | | | | | |
|--------------------|-------------------|--------|--------------|---|--------------------------------------|---|
| Calibration Point | Manometer Reading | | | Q _{std} (m ³ / min.) X-axis | Continuous Flow Recorder, W (CFM) | IC (W(P _a /1013.3x298/T _a) ^{1/2} /35.31) Y-axis |
| | (up) | (down) | (difference) | | | |
| 1 | 6.2 | 6.2 | 12.4 | 1.7419 | 60 | 59.4747 |
| 2 | 5.3 | 5.3 | 10.6 | 1.6115 | 52 | 51.5448 |
| 3 | 4.1 | 4.1 | 8.2 | 1.4190 | 43 | 42.6236 |
| 4 | 2.7 | 2.7 | 5.4 | 1.1541 | 24 | 23.7899 |
| 5 | 1.4 | 1.4 | 2.8 | 0.8349 | 12 | 11.8949 |

By Linear Regression of Y on X

Slope, m = 53.7477 Intercept, b = -34.8156
 Correlation Coefficient* = 0.9945
 Calibration Accepted = Yes/No**

* if Correlation Coefficient < 0.990, check and recalibration again.

** Delete as appropriate.

Remarks : _____

Calibrated by : Felix Li Checked by : Pauline Wong
 Date : 2-Jul-14 Date : 2-Jul-14



Lam Geotechnics Limited

Calibration Data for High Volume Sampler (TSP Sampler)

Location : CMA5a Calibration Date : 21-Jun-14
 Equipment no. : EL380 Calibration Due Date : 21-Aug-14

CALIBRATION OF CONTINUOUS FLOW RECORDER

| Ambient Condition | | | |
|-----------------------------|-----|--------|--------------------------|
| Temperature, T _a | 301 | Kelvin | Pressure, P _a |
| | | | 1003 mmHg |

| Orifice Transfer Standard Information | | | |
|---------------------------------------|-----------|---|----------|
| Equipment No. | EL086 | Slope, m _c | 2.01968 |
| | | Intercept, b _c | -0.02746 |
| Last Calibration Date | 15-Jul-13 | $(H \times P_a / 1013.3 \times 298 / T_a)^{1/2}$ $= m_c \times Q_{std} + b_c$ | |
| Next Calibration Date | 15-Jul-14 | | |

| Calibration of TSP | | | | | | |
|--------------------|-------------------|--------|--------------|---|--------------------------------------|---|
| Calibration Point | Manometer Reading | | | Q _{std} (m ³ / min.) X-axis | Continuous Flow Recorder, W (CFM) | IC (W(P _a /1013.3x298/T _a) ^{1/2} /35.31) Y-axis |
| | (up) | (down) | (difference) | | | |
| 1 | 5.6 | 5.6 | 11.2 | 1.6539 | 61 | 60.3860 |
| 2 | 4.8 | 4.8 | 9.6 | 1.5323 | 52 | 51.4766 |
| 3 | 3.5 | 3.5 | 7.0 | 1.3104 | 42 | 41.5772 |
| 4 | 2.3 | 2.3 | 4.6 | 1.0648 | 26 | 25.7383 |
| 5 | 1.2 | 1.2 | 2.4 | 0.7729 | 13 | 12.8691 |

By Linear Regression of Y on X

Slope, m = 53.8279 Intercept, b = -29.7835
 Correlation Coefficient* = 0.9974
 Calibration Accepted = Yes/No**

* if Correlation Coefficient < 0.990, check and recalibration again.

** Delete as appropriate.

Remarks : _____

Calibrated by : Felix Li Checked by : Pauline Wong
 Date : 21-Jun-14 Date : 21-Jun-14



Lam Geotechnics Limited

Calibration Data for High Volume Sampler (TSP Sampler)

Location : CMA6a Calibration Date : 13-May-14
 Equipment no. : EL448 Calibration Due Date : 13-Jul-14

CALIBRATION OF CONTINUOUS FLOW RECORDER

| Ambient Condition | | | |
|-----------------------------|-----|--------|--------------------------|
| Temperature, T _a | 300 | Kelvin | Pressure, P _a |
| | | | 1007 mmHg |

| Orifice Transfer Standard Information | | | |
|---------------------------------------|-----------|---|----------|
| Equipment No. | EL086 | Slope, m _c | 2.01968 |
| | | Intercept, b _c | -0.02746 |
| Last Calibration Date | 15-Jul-13 | $(H \times P_a / 1013.3 \times 298 / T_a)^{1/2}$ $= m_c \times Q_{std} + b_c$ | |
| Next Calibration Date | 15-Jul-14 | | |

| Calibration of TSP | | | | | | |
|--------------------|-------------------|--------|--------------|---|--------------------------------------|--|
| Calibration Point | Manometer Reading | | | Q _{std} (m ³ / min.) X-axis | Continuous Flow Recorder, W (CFM) | IC $(W(P_a/1013.3 \times 298/T_a)^{1/2}/35.31)$ Y-axis |
| | (up) | (down) | (difference) | | | |
| 1 | 6.1 | 6.1 | 12.2 | 1.7319 | 62 | 61.6006 |
| 2 | 5.0 | 5.0 | 10.0 | 1.5692 | 52 | 51.6650 |
| 3 | 4.0 | 4.0 | 8.0 | 1.4050 | 42 | 41.7294 |
| 4 | 2.4 | 2.4 | 4.8 | 1.0914 | 25 | 24.8389 |
| 5 | 1.5 | 1.5 | 3.0 | 0.8657 | 13 | 12.9163 |

By Linear Regression of Y on X

Slope, m = 55.9776 Intercept, b = -36.0474
 Correlation Coefficient* = 0.9995
 Calibration Accepted = Yes/No**

* if Correlation Coefficient < 0.990, check and recalibration again.

** Delete as appropriate.

Remarks : _____

Calibrated by : Felix Li Checked by : Derek Lo
 Date : 13-May-14 Date : 13-May-14



Lam Geotechnics Limited

Calibration Data for High Volume Sampler (TSP Sampler)

Location : CMA6a
 Equipment no. : EL448

Calibration Date : 2-Jul-14
 Calibration Due Date : 2-Sep-14

CALIBRATION OF CONTINUOUS FLOW RECORDER

| Ambient Condition | | | |
|-----------------------------|-----|--------|--------------------------|
| Temperature, T _a | 302 | Kelvin | Pressure, P _a |
| | | | 1009 mmHg |

| Orifice Transfer Standard Information | | | |
|---------------------------------------|-----------|--|----------|
| Equipment No. | EL086 | Slope, m _c | 2.01968 |
| | | Intercept, b _c | -0.02746 |
| Last Calibration Date | 15-Jul-13 | $\left(H \times P_a / 1013.3 \times 298 / T_a \right)^{1/2}$ $= m_c \times Q_{std} + b_c$ | |
| Next Calibration Date | 15-Jul-14 | | |

| Calibration of TSP | | | | | | |
|--------------------|-------------------|--------|--------------|---|--------------------------------------|---|
| Calibration Point | Manometer Reading | | | Q _{std} (m ³ / min.) X-axis | Continuous Flow Recorder, W (CFM) | IC (W(P _a /1013.3x298/T _a) ^{1/2} /35.31) Y-axis |
| | (up) | (down) | (difference) | | | |
| 1 | 6.2 | 6.2 | 12.4 | 1.7419 | 60 | 59.4747 |
| 2 | 5.1 | 5.1 | 10.2 | 1.5811 | 52 | 51.5448 |
| 3 | 4.3 | 4.3 | 8.6 | 1.4529 | 42 | 41.6323 |
| 4 | 2.3 | 2.3 | 4.6 | 1.0662 | 23 | 22.7986 |
| 5 | 1.6 | 1.6 | 3.2 | 0.8916 | 12 | 11.8949 |

By Linear Regression of Y on X

Slope, m = 55.4756 Intercept, b = -37.2410
 Correlation Coefficient* = 0.9984
 Calibration Accepted = Yes/No**

* if Correlation Coefficient < 0.990, check and recalibration again.

** Delete as appropriate.

Remarks : _____

Calibrated by : Henry Lau
 Date : 2-Jul-14

Checked by : Pauline Wong
 Date : 2-Jul-14



Lam Geotechnics Limited

Calibration Data for High Volume Sampler (TSP Sampler)

Location : MA1e Calibration Date : 13-May-14
 Equipment no. : EL455 Calibration Due Date : 13-Jul-14

CALIBRATION OF CONTINUOUS FLOW RECORDER

| Ambient Condition | | | | | | |
|---------------------------------------|-------------------|---|--------------------------|---|--------------------------------------|---|
| Temperature, T _a | 300 | Kelvin | Pressure, P _a | 1007 | mmHg | |
| Orifice Transfer Standard Information | | | | | | |
| Equipment No. | EL086 | Slope, m _c | 2.01968 | Intercept, b _c | -0.02746 | |
| Last Calibration Date | 15-Jul-13 | $(H \times P_a / 1013.3 \times 298 / T_a)^{1/2}$ $= m_c \times Q_{std} + b_c$ | | | | |
| Next Calibration Date | 15-Jul-14 | | | | | |
| Calibration of TSP | | | | | | |
| Calibration Point | Manometer Reading | | | Q _{std} (m ³ / min.) X-axis | Continuous Flow Recorder, W (CFM) | IC (W(P _a /1013.3x298/T _a) ^{1/2} /35.31) Y-axis |
| | (up) | (down) | (difference) | | | |
| 1 | 6.2 | 6.2 | 12.4 | 1.7459 | 62 | 61.6006 |
| 2 | 5.1 | 5.1 | 10.2 | 1.5847 | 52 | 51.6650 |
| 3 | 4.1 | 4.1 | 8.2 | 1.4223 | 43 | 42.7230 |
| 4 | 2.5 | 2.5 | 5.0 | 1.1136 | 26 | 25.8325 |
| 5 | 1.6 | 1.6 | 3.2 | 0.8936 | 15 | 14.9034 |
| By Linear Regression of Y on X | | | | | | |
| Slope, m | | = | 54.6458 | Intercept, b | | = -34.5374 |
| Correlation Coefficient* | | = | 0.9995 | | | |
| Calibration Accepted | | = | Yes/No** | | | |

* if Correlation Coefficient < 0.990, check and recalibration again.

** Delete as appropriate.

Remarks : _____

Calibrated by : Felix Li Checked by : Derek Lo
 Date : 13-May-14 Date : 13-May-14



Lam Geotechnics Limited

Calibration Data for High Volume Sampler (TSP Sampler)

Location : MA1e
 Equipment no. : EL455

Calibration Date : 21-Jun-14
 Calibration Due Date : 21-Aug-14

CALIBRATION OF CONTINUOUS FLOW RECORDER

| Ambient Condition | | | |
|-----------------------------|-----|--------|--------------------------|
| Temperature, T _a | 301 | Kelvin | Pressure, P _a |
| | | | 1003 mmHg |

| Orifice Transfer Standard Information | | | |
|---------------------------------------|-----------|--|----------|
| Equipment No. | EL086 | Slope, m _c | 2.01968 |
| | | Intercept, b _c | -0.02746 |
| Last Calibration Date | 15-Jul-13 | $\left(H \times P_a / 1013.3 \times 298 / T_a \right)^{1/2}$ $= m_c \times Q_{std} + b_c$ | |
| Next Calibration Date | 15-Jul-14 | | |

| Calibration of TSP | | | | | | |
|--------------------|-------------------|--------|--------------|---|--------------------------------------|---|
| Calibration Point | Manometer Reading | | | Q _{std} (m ³ / min.) X-axis | Continuous Flow Recorder, W (CFM) | IC (W(P _a /1013.3x298/T _a) ^{1/2} /35.31) Y-axis |
| | (up) | (down) | (difference) | | | |
| 1 | 6.5 | 6.5 | 13.0 | 1.7808 | 57 | 56.4262 |
| 2 | 4.2 | 4.2 | 8.4 | 1.4342 | 48 | 47.5168 |
| 3 | 3.3 | 3.3 | 6.6 | 1.2728 | 42 | 41.5772 |
| 4 | 2.1 | 2.1 | 4.2 | 1.0181 | 32 | 31.6779 |
| 5 | 1.4 | 1.4 | 2.8 | 0.8338 | 28 | 27.7182 |

By Linear Regression of Y on X

Slope, m = 31.5589 Intercept, b = 0.9687
 Correlation Coefficient* = 0.9958
 Calibration Accepted = Yes/No**

* if Correlation Coefficient < 0.990, check and recalibration again.

** Delete as appropriate.

Remarks : _____

Calibrated by : Felix Li
 Date : 21-Jun-14

Checked by : Pauline Wong
 Date : 21-Jun-14



Lam Geotechnics Limited

Calibration Data for High Volume Sampler (TSP Sampler)

Location : MA1w
 Equipment no. : EL080

Calibration Date : 21-Jun-14
 Calibration Due Date : 21-Aug-14

CALIBRATION OF CONTINUOUS FLOW RECORDER

| Ambient Condition | | | |
|-----------------------------|-----|--------|--------------------------|
| Temperature, T _a | 301 | Kelvin | Pressure, P _a |
| | | | 1003 mmHg |

| Orifice Transfer Standard Information | | | |
|---------------------------------------|-----------|--|----------|
| Equipment No. | EL086 | Slope, m _c | 2.01968 |
| | | Intercept, b _c | -0.02746 |
| Last Calibration Date | 15-Jul-13 | $\left(H \times P_a / 1013.3 \times 298 / T_a \right)^{1/2}$ $= m_c \times Q_{std} + b_c$ | |
| Next Calibration Date | 15-Jul-14 | | |

| Calibration of TSP | | | | | | |
|--------------------|-------------------|--------|--------------|---|--------------------------------------|---|
| Calibration Point | Manometer Reading | | | Q _{std} (m ³ / min.) X-axis | Continuous Flow Recorder, W (CFM) | IC (W(P _a /1013.3x298/T _a) ^{1/2} /35.31) Y-axis |
| | (up) | (down) | (difference) | | | |
| 1 | 6.3 | 6.3 | 12.6 | 1.7534 | 53 | 52.4665 |
| 2 | 5.1 | 5.1 | 10.2 | 1.5790 | 48 | 47.5168 |
| 3 | 4.2 | 4.2 | 8.4 | 1.4342 | 40 | 39.5974 |
| 4 | 2.6 | 2.6 | 5.2 | 1.1313 | 30 | 29.6980 |
| 5 | 1.8 | 1.8 | 3.6 | 0.9436 | 24 | 23.7584 |

By Linear Regression of Y on X

Slope, m = 36.2029 Intercept, b = -10.9288
 Correlation Coefficient* = 0.9965
 Calibration Accepted = Yes/No**

* if Correlation Coefficient < 0.990, check and recalibration again.

** Delete as appropriate.

Remarks : _____

Calibrated by : Felix Li
 Date : 21-Jun-14

Checked by : Pauline Wong
 Date : 21-Jun-14



Lam Geotechnics Limited

Calibration Data for High Volume Sampler (TSP Sampler)

Location : ACL1 Calibration Date : 21-Jun-14
 Equipment no. : EL222 Calibration Due Dat : 21-Aug-14

CALIBRATION OF CONTINUOUS FLOW RECORDER

| Ambient Condition | | | |
|-----------------------------|-----|--------|--------------------------|
| Temperature, T _a | 301 | Kelvin | Pressure, P _a |
| | | | 1003 mmHg |

| Orifice Transfer Standard Information | | | | | |
|---------------------------------------|-----------|---|---------|---------------------------|----------|
| Equipment No. | EL086 | Slope, m _c | 2.01968 | Intercept, b _c | -0.02746 |
| Last Calibration Date | 15-Jul-13 | $(H \times P_a / 1013.3 \times 298 / T_a)^{1/2}$ $= m_c \times Q_{std} + b_c$ | | | |
| Next Calibration Date | 15-Jul-14 | | | | |

| Calibration of TSP | | | | | | |
|--------------------|-------------------|--------|--------------|---|--------------------------------------|---|
| Calibration Point | Manometer Reading | | | Q _{std} (m ³ / min.) X-axis | Continuous Flow Recorder, W (CFM) | IC (W(P _a /1013.3x298/T _a) ^{1/2} /35.31) Y-axis |
| | (up) | (down) | (difference) | | | |
| 1 | 5.8 | 5.8 | 11.6 | 1.6830 | 56 | 55.4363 |
| 2 | 4.4 | 4.4 | 8.8 | 1.4676 | 48 | 47.5168 |
| 3 | 3.7 | 3.7 | 7.4 | 1.3469 | 43 | 42.5672 |
| 4 | 2.2 | 2.2 | 4.4 | 1.0417 | 34 | 33.6578 |
| 5 | 1.5 | 1.5 | 3.0 | 0.8626 | 26 | 25.7383 |

By Linear Regression of Y on X

Slope, m = 35.2472 Intercept, b = -4.1457
 Correlation Coefficient* = 0.9981
 Calibration Accepted = Yes/No**

* if Correlation Coefficient < 0.990, check and recalibration again.

** Delete as appropriate.

Remarks : _____

Calibrated by : Felix Li Checked by : Pauline Wong
 Date : 21-Jun-14 Date : 21-Jun-14



Lam Geotechnics Limited

Calibration Data for High Volume Sampler (TSP Sampler)

Location : ACL2a Calibration Date : 3-Jun-14
 Equipment no. : EL111 Calibration Due Dat : 3-Aug-14

CALIBRATION OF CONTINUOUS FLOW RECORDER

| Ambient Condition | | | |
|-----------------------------|-----|--------|--------------------------|
| Temperature, T _a | 303 | Kelvin | Pressure, P _a |
| | | | 1004 mmHg |

| Orifice Transfer Standard Information | | | | | |
|---------------------------------------|-----------|---|---------|---------------------------|----------|
| Equipment No. | EL086 | Slope, m _c | 2.01968 | Intercept, b _c | -0.02746 |
| Last Calibration Date | 15-Jul-13 | $(H \times P_a / 1013.3 \times 298 / T_a)^{1/2}$ $= m_c \times Q_{std} + b_c$ | | | |
| Next Calibration Date | 15-Jul-14 | | | | |

| Calibration of TSP | | | | | | |
|--------------------|-------------------|--------|--------------|---|--------------------------------------|---|
| Calibration Point | Manometer Reading | | | Q _{std} (m ³ / min.) X-axis | Continuous Flow Recorder, W (CFM) | IC (W(P _a /1013.3x298/T _a) ^{1/2} /35.31) Y-axis |
| | (up) | (down) | (difference) | | | |
| 1 | 6.2 | 6.2 | 12.4 | 1.7347 | 61 | 60.2164 |
| 2 | 5.1 | 5.1 | 10.2 | 1.5746 | 53 | 52.3191 |
| 3 | 4.2 | 4.2 | 8.4 | 1.4302 | 46 | 45.4091 |
| 4 | 2.6 | 2.6 | 5.2 | 1.1282 | 33 | 32.5761 |
| 5 | 1.6 | 1.6 | 3.2 | 0.8879 | 23 | 22.7045 |

By Linear Regression of Y on X

Slope, m = 44.0592 Intercept, b = -16.8841

Correlation Coefficient* = 0.9993

Calibration Accepted = Yes/No**

* if Correlation Coefficient < 0.990, check and recalibration again.

** Delete as appropriate.

Remarks : _____

Calibrated by : Henry Checked by : Derek Lo
 Date : 3-Jun-14 Date : 3-Jun-14